



CGA P-83—2019

GUIDELINES FOR CLEANING EXTERNALLY CONTAMINATED MEDICAL GAS CONTAINERS

FIRST EDITION

This document licensed for the sole use of the purchaser. It may not be shared with any other person or used after the expiration date without the express written permission of the CGA. Any unauthorized use, reproduction, distribution, or modification of this printed page will result in a \$5,000 liquidated damages fee plus loss of access to CGA publications for one year for your company.

PREFACE

As part of a program of harmonization of industry standards, the Compressed Gas Association (CGA) has issued CGA P-83, *Guidelines for Cleaning Externally Contaminated Medical Gas Containers*, jointly produced by members of the International Harmonization Council and originally published by the European Industrial Gases Association (EIGA) as EIGA Doc 222, *Guidelines for Cleaning Externally Contaminated Medical Gas Containers*.

This publication is intended as an international harmonized standard for the worldwide use and application of all members of the Asia Industrial Gases Association (AIGA), Compressed Gas Association (CGA), European Industrial Gases Association (EIGA), and Japan Industrial and Medical Gases Association (JIMGA). Each association's technical content is identical, except for regional regulatory requirements and minor changes in formatting and spelling.

PLEASE NOTE:

The information contained in this document was obtained from sources believed to be reliable and is based on technical information and experience currently available from members of the Compressed Gas Association, Inc. and others. However, the Association or its members, jointly or severally, make no guarantee of the results and assume no liability or responsibility in connection with the information or suggestions herein contained. Moreover, it should not be assumed that every acceptable commodity grade, test or safety procedure or method, precaution, equipment or device is contained within, or that abnormal or unusual circumstances may not warrant or suggest further requirements or additional procedure.

This document is subject to periodic review, and users are cautioned to obtain the latest edition. The Association invites comments and suggestions for consideration. In connection with such review, any such comments or suggestions will be fully reviewed by the Association after giving the party, upon request, a reasonable opportunity to be heard. Proposed changes may be submitted via the Internet at our web site, www.cganet.com.

This document should not be confused with federal, state, provincial, or municipal specifications or regulations; insurance requirements; or national safety codes. While the Association recommends reference to or use of this document by government agencies and others, this document is purely voluntary and not binding unless adopted by reference in regulations.

A listing of all publications, audiovisual programs, safety and technical bulletins, and safety posters is available via the Internet at our website at www.cganet.com. For more information contact CGA at Phone: 703-788-2700, ext. 799. E-mail: customerservice@cganet.com.

Work Item 15-043
Safety and Health Committee

FIRST EDITION: 2019

2019 © – Reproduced with permission from the European Industrial Gases Association. All rights reserved.

Contents	Page
1 Introduction.....	1
2 Scope and purpose	1
3 Definitions.....	1
4 Principles for safe handling of externally contaminated medical gas containers	2
5 Recommended cleaning and disinfection methods.....	2
6 Procedures and training	4
7 References	4

1 Introduction

Medical gas containers can become externally contaminated due to their use in healthcare facilities and homecare settings. They can also become contaminated as a result of natural disasters such as floods, hurricanes, fires, or earthquakes.

It is important that these containers are cleaned in a manner that removes the contaminants, does not cause damage to the container and its accessories, and does not lead to a hazardous situation during the use of the container's contents.

2 Scope and purpose

2.1 Scope

This publication addresses the handling and external condition of medical gas containers and the removal of contaminants from these containers.

The internal condition or product quality of the container is not addressed in this publication.

2.2 Purpose

The purpose of this publication is to ensure that external contamination is removed from medical gas containers so that the process does not damage the container or its accessories and protects personnel handling medical gas containers.

3 Definitions

For the purpose of this publication, the following definitions apply.

3.1 Publication terminology

3.1.1 Shall

Indicates that the procedure is mandatory. It is used wherever the criterion for conformance to specific recommendations allows no deviation.

3.1.2 Should

Indicates that a procedure is recommended.

3.1.3 May

Indicates that the procedure is optional.

3.1.4 Will

Is used only to indicate the future, not a degree of requirement.

3.1.5 Can

Indicates a possibility or ability.

3.2 Technical definitions

3.2.1 Cleaning

Removal of contaminants to the extent necessary for further processing or for intended use.

3.2.2 Disinfection

Process to destroy viable microorganisms.

3.2.3 External contaminant

Foreign matter not intentionally added that is capable of producing an adverse effect on the medical gas container or human health.

NOTE—Examples of foreign matter include lubricants, body fluids, and other chemical, physical, and biological contaminants.

3.2.4 Healthcare facility

Hospital, clinic, or similar facility that provides patients with their healthcare needs.

NOTE—Healthcare facility can include emergency services such as ambulance companies.

3.2.5 Medical gas container

Gas cylinder, bundle of cylinders, or a cryogenic container including their accessories.

4 Principles for safe handling of externally contaminated medical gas containers

4.1 General

Medical gas containers that could have external contaminants are likely to be first handled at the customer premises. As this is the first point of contact, gas suppliers shall have a procedure in place for the handling of externally contaminated medical gas containers. This procedure shall address the requirements for personal protective equipment (PPE) and other regulatory requirements. See CGA P-44, *Selection of Personal Protective Equipment*, for information on PPE requirements [1].¹

Personnel involved in handling medical gas containers shall be trained in this procedure, see Section 6.

Medical gas containers that are externally contaminated shall be quarantined until cleaned.

4.2 Customer guidance

The gas supplier shall advise their customers that:

- They have a duty to return medical gas containers that are free from contamination to the gas supplier; and
- When contamination by infectious pathogens is suspected, they shall take measures to reduce the risks to themselves and to gas supplier employees.

The customer shall review the guidelines provided by the gas supplier for cleaning medical gas containers. If additional information is needed, the customer shall contact the gas supplier.

5 Recommended cleaning and disinfection methods

5.1 General

The cleaning and disinfection procedure should be performed at sites in designated areas with appropriate equipment and facilities and using a safe and efficient waste disposal system in accordance with local regulations.

Cleaning normally is accomplished manually or mechanically using potable water with detergents or enzymatic products. Thorough cleaning is essential before disinfection because inorganic and organic materials that remain on the surface of the container interfere with the effectiveness of this process.

The cleaning method shall not have a detrimental effect on the medical gas container, valve, and other accessories such as regulators and flowmeters. Cleaning materials shall be approved by the gas supplier prior to use.

¹ References are shown by bracketed numbers and are listed in order of appearance in the reference section.

When cleaning untreated metallic parts of a medical gas container, avoid using cleaning agents that contain ammonia, amine-based compounds, or chlorine-based compounds (such as bleach) as they can cause corrosion of steel or aluminum-alloy components or stress corrosion cracking (SCC) of brass including copper-alloy components.

For additional information regarding the effects of ammonia and SCC, see CGA V-12, *Leak Detection Fluids Cylinder Packages* [2].

When cleaning is performed at locations other than the gas supplier, labels on the medical gas container shall not be damaged or removed during the cleaning process.

The healthcare facility should contact the gas supplier if there is any doubt concerning the method to be used to clean the container. They should report whether they have cleaned the medical gas container so that the gas supplier can verify the removal of any external contamination prior to refilling.

If the container has been contaminated or is suspected to have been contaminated, the following options are available to the healthcare facility:

- clean it prior to returning to the gas supplier, see 5.2 and 5.3; or
- if cleaning at the healthcare facility is not possible, at a minimum, contaminated medical gas containers should be identified and segregated, for example, by covering with a plastic bag that is labelled with the nature of the hazard and noting that it requires cleaning.

In the second case, the healthcare facility shall contact the gas supplier for further guidance.

5.2 Initial cleaning

For the initial cleaning, the preferred method is to use potable hot water not exceeding 122 °F (50 °C) to first remove foreign matter. Valve outlets and inlets shall be closed and covered to prevent ingress of water. The valves shall be protected unless they are externally contaminated.

The medical gas container shall not be immersed in water.

5.3 Disinfection

Factors that affect the efficacy of disinfection include:

- prior cleaning of the medical gas container;
- type and level of microbial contamination;
- concentration of and exposure time to the disinfection agent;
- physical nature of the object (for example, crevices); and
- temperature and pH of the disinfection process.

After the contamination has been removed, the medical gas container shall be disinfected by using, for example, isopropyl alcohol (IPA) or equivalent disinfectant wipes. Disinfection agents shall be approved by the gas supplier and used in accordance with the manufacturer's recommendations.

The application of alcohol-based disinfectants shall be limited to prevent excessive amounts on the valve and/or cylinder that could cause a potential fire risk.

Ensure that all residual disinfection agents are removed from the medical gas container.

6 Procedures and training

Healthcare facility, gas supplier, and third-party personnel who handle medical gas containers shall receive training and instructions on how to handle contaminated medical gas containers. This shall include using the appropriate PPE, possibly covering the container with a plastic bag, and adding a warning label. Whenever handling contaminated medical gas containers, personnel shall dispose of gloves, contaminated clothing, and any other contaminated materials as hazardous waste in accordance with local regulations. Personnel shall wash their hands after handling contaminated medical gas containers.

Operational management shall define specific written instructions for their personnel.

7 References

Unless otherwise specified, the latest edition shall apply.

[1] CGA P-44, *Selection of Personal Protective Equipment*, Compressed Gas Association, Inc.
www.cganet.com

NOTE—This publication is part of an international program for industry standards. The technical content of each regional document is identical, except for regional regulatory requirements. See the referenced document preface for a list of harmonized regional references.

[2] CGA V-12, *Leak Detection Fluids Use with Gas Cylinder Packages*, Compressed Gas Association, Inc.
www.cganet.com

NOTE—This publication is part of an international program for industry standards. The technical content of each regional document is identical, except for regional regulatory requirements. See the referenced document preface for a list of harmonized regional references.

CGA

Compressed Gas Association

The Standard For Safety Since 1913

8484 Westpark Drive
Suite 220
McLean, VA 22102
USA

703.788.2700

www.cganet.com